



## A.H. Systems, Inc.

9710 Cozycroft Ave.  
Chatsworth, CA 91311



Tel: (818) 998-0223



sales@AHSystems.com

Fax: (818) 998-6892



www.AHSystems.com

## BCP-611

### Broadband Current Probe

10 KHz – 150 MHz

This probe is capable of measuring pulse currents up to 100 amperes.



Frequency Range : 10 KHz - 150 MHz

Transfer Impedance : -25 to +5 dBΩ

Max Primary Current : 800 Amps (DC to 60 Hz)

Max Primary Current : 450 Amps (400 Hz)

Max Primary Current CW : 4 Amps

Max Pulse Current : 100 Amps

Connector : BNC-Type, Female

### **Physical Dimensions**

Aperture: 1.25" (33 mm)

Diameter: 4.5" (114 mm)

Weight: 1.4 lb.'s (0.65 Kg)

### **Features**

- Measure conducted emissions from 10 kHz to 150 MHz
- Individually Calibrated (Transfer Impedance calibration included)
- High Current Capability
- Split Type Clamp-on Design
- Three Year Warranty

Conducted currents can be measured without making direct contact with the source conductor or metallic surface by means of clamp-on current probes. The BCP-611 Current Probe is designed to permit field intensity meters, spectrum analyzers, and other 50 ohm impedance instruments to measure quantitative magnitudes of current. Measurements can be made on single and multi-conductor cables, ground and bonding straps, shielded conduits and on coaxial cables.

For ease and convenience of performing conducted measurements, all of our current probes utilize the split type clamp-on design. Small and lightweight, each Current Probe is manufactured to exacting standards, thus insuring repeatable performance.

### **Recommended Accessories**

- CPF-630 Current Probe Fixture
- SAC-212 BNC/N Cable, 3 Meter



**A.H. Systems, Inc.**

9710 Cozycroft Ave.  
Chatsworth, CA 91311



Tel: (818) 998-0223  
Fax: (818) 998-6892

◆ sales@AHSystems.com  
◆ www.AHSystems.com



**A.H. Systems, Inc.**  
9710 Cozycroft Ave.  
Chatsworth, CA 91311  
818.998.0223 fax 818.998.6892  
sales@AHSystems.com www.AHSystems.com



**Broadband  
Current Probe  
Calibration  
Model: BCP-611**

Transfer Impedance Conversion Formula:

$$\text{dB}\mu\text{A} = \text{dB}\mu\text{V} - \text{dB}\Omega + \text{cable loss}$$

